

Republic of the Philippines  
University of the Philippines Mindanao

Date: 12 February 2018

**ADDENDUM NO. 01-2018**

Name of Contract: **Supply of Laboratory Equipment-Laminar Flow Hood, Sterilizer, Automated Dispenser, pH Meter, Magnetic Stirrer, Top Loading Balance and Water Distillation System**

This Addendum is issued for the revised technical specifications of the following items:

Description	Quantity
<b>LAMINAR FLOW HOOD</b> Sterile Chamber External Dimension: 1800x820x2010mm Power Supply: 220V 50/60Hz Net Weight: 250kg Input Power: 1310W Horizontal Fabricated	2 units
<b>AUTOMATED DISPENSER</b> For dispensing of culture media, agar or any liquid diluent Single Pump Head with 3 rollers Dose volume: 50ul to 99L Flow rate up to 1.25L/min Pump Head rotating speed: 10 to 550 RPM Accepted tube inner diameter (ID): from 1 to 8 mm Accepted tube wall thickness: 1.6 mm only Power: 100-240V 50-60Hz Machine Size: (WxDxH): 17x28x21 cm, weight 5.3kg Box size: (WxDxH): 50x40x40cm, weight 9kg	2 units

For guidance and information of all concerned.

  
**PROF. VICENTE B. CALAG**  
Vice Chair  
Bids and Awards Committee

Received by:

\_\_\_\_\_  
Signature over Printed Name/Name of Company